

Wir

**B. Braun Melsungen AG  
Carl-Braun-Straße 1  
34212 Melsungen  
Deutschland/Germany**

We

erklären in eigener Verantwortung,  
dass das/die Produkt/e**Redyrob® COMP,  
Redyrob® Trans Plus,  
Schwerkraft-Drainagebeutel 1,5L**Geschlossenes Wunddrainagesysteme für  
Anwendungen

(Artikelnummern siehe Anlage I)

mit den Anforderungen der folgenden Richtlinie  
übereinstimmt/übereinstimmenRichtlinie 93/42/EWG des Rates vom 14. Juni  
1993über Medizinprodukte  
geändert durch Richtlinie 2007/47/EG**Konformitätsbewertungsverfahren**  
nach Anhang II (ausgenommen Abschnitt 4)  
nach Anhang VII und V  
der oben genannten Richtlinie**Klassifizierung**  
gemäß Anhang IX der  
oben genannten Richtlinie  
Klasse IIa  
Klasse I steril**Benannte Stelle**  
TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München  
Deutschland  
Kennnummer 0123**Datum der ersten CE-Kennzeichnung**  
1994-12**Gültig bis**  
2024-05-26hereby declare in our own responsibility  
that the product/s**Redyrob® COMP,  
Redyrob® Trans Plus,  
Schwerkraft-Drainagebeutel 1,5L**Closed wound drainage system for each  
indication

(article numbers see attachment I)

is/are in compliance with the following directive

Council Directive 93/42/EEC of 14<sup>th</sup> June 1993  
concerning Medical Devices  
amended by Directive 2007/47/EC**Conformity Assessment Procedure**  
according to annex II (excluding section 4)  
according to annex VII and V  
of the Council Directive named above**Classification**  
according to annex IX of the  
Council Directive named above  
Class IIa  
Class I sterile**Notified Body**  
TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München  
Germany  
Identification number 0123**Date of first CE-marking**  
1994-12**Valid until**  
2024-05-26

**Anlage I / Attachment I**

<b>Art.-Nr. / Art. No.</b>	<b>Produktname / Product name</b>	<b>Klasse / Class</b>
5526000	Redyrob® COMP	Ila
5526159	Schwerkraft-Drainagebeutel 1,5L	I steril / I sterile
5526604	Redyrob® Trans Plus	Ila
5526680	Redyrob® COMP	Ila

**History**

<b>Version</b>	<b>Description of the changes</b>
09	Update TÜV address
10	Change classification of art. no. 5526159 from IIa to I sterile

Title: Declaration of Conformity - 39.05.300 - Redyrob Initiator: Caroline ? Herbst

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Herbst, Caroline (fuehcade)  
Title: Administrator Regulatory Affairs  
Date: Tuesday, 21 May 2019, 14:20 W. Europe Daylight Time  
Meaning: Document signed as Author  
=====

UserName: Brand, Thomas (brantode)  
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices  
Date: Tuesday, 21 May 2019, 18:49 W. Europe Daylight Time  
Meaning: Approve Document  
=====

UserName: Seidel, Stefan (seidstde)  
Title: HC-RA-DE08E - Head of Regulatory Affairs CoE IV Systems  
Date: Wednesday, 22 May 2019, 09:49 W. Europe Daylight Time  
Meaning: Approve Document  
=====